

Spectral Gamma-Ray Borehole Log Data Report

Page 1 of 2

Log Event A

Borehole 30-05-10

Borehole Information

N-Coord: 42,893 W-Coord: 48,366 TOC Elevation: 646.23

Water Level, ft : Date Drilled : <u>11/30/1972</u>

Casing Record

Type: Steel-welded Thickness: 0.280 ID, in.: 6

Top Depth, ft. : $\underline{0}$ Bottom Depth, ft. : $\underline{135}$

Borehole Notes:

This borehole was drilled in November 1972 to a depth of 135 ft using 6-in. casing. The drilling report does not indicate if the borehole casing was perforated or grouted. The casing thickness is presumed to be 0.280 in., on the basis of the published thickness for schedule-40, 6-in. steel tubing. The top of the casing, which is the zero reference for the SGLS, is approximately flush with the ground surface. The current total depth of the borehole was measured at 135.8 ft below the top of the casing using a weighted tape, indicating that the borehole casing was probably extended to accommodate fill material that was added to the surface of the tank farm after the borehole was completed.

Equipment Information

 Logging System :
 2
 Detector Type :
 HPGe
 Detector Efficiency:
 35.0 %

 Calibration Date :
 10/1996
 Calibration Reference :
 GJO-HAN-13
 Logging Procedure : P-GJPO-1783

Log Run Information

 Log Run Number :
 1
 Log Run Date :
 01/15/1997
 Logging Engineer:
 Bob Spatz

 Start Depth, ft.:
 0.0
 Counting Time, sec.:
 100
 L/R : I
 Shield : N

Finish Depth, ft.: 0.0 Counting Time, sec.: 100 L/R: L Shield: N Finish Depth, ft.: 11.0 MSA Interval, ft.: 0.5 Log Speed, ft/min.: 100

Log Run Number: 2 Log Run Date: 01/16/1997 Logging Engineer: Bob Spatz

Start Depth, ft.: $\underline{10.0}$ Counting Time, sec.: $\underline{100}$ L/R: \underline{L} Shield: \underline{N} Finish Depth, ft.: $\underline{91.0}$ MSA Interval, ft.: $\underline{0.5}$ Log Speed, ft/min.: \underline{n}/a

Log Run Number: 3 Log Run Date: 01/17/1997 Logging Engineer: Bob Spatz

Start Depth, ft.: $\underline{135.5}$ Counting Time, sec.: $\underline{100}$ L/R: \underline{L} Shield: \underline{N} Finish Depth, ft.: $\underline{90.0}$ MSA Interval, ft.: $\underline{0.5}$ Log Speed, ft/min.: $\underline{n/a}$



Spectral Gamma-Ray Borehole Log Data Report

Page 2 of 2

Log Event A

Borehole 30-05-10

Analysis Information

Analyst: E. Larsen

Data Processing Reference : P-GJPO-1787 Analysis Date : 07/02/1997

Analysis Notes:

This borehole was logged by the SGLS in three log runs. The pre- and post-survey field verification spectra met the acceptance criteria established for the peak shape and detector efficiency, confirming that the SGLS was operating within specifications. The energy calibration and peak-shape calibration from these spectra were used to establish the channel-to-energy parameters used in processing the spectra acquired during the logging operation.

Casing correction factors for a 0.280-in.-thick steel casing were applied during analysis.

The man-made radionuclides Cs-137, Co-60, and U-235 were detected in this borehole. Continuous Cs-137 contamination was detected from the ground surface to 16 ft, 26 to 30 ft, 114.5 to 116.5 ft, and at the bottom of the logged interval (134.5 to 135.5 ft). Isolated minor occurrences of Cs-137 contamination are scattered throughout the rest of the borehole. Several small zones of Co-60 contamination were detected between 13 and 53.5 ft. A single occurrence of U-235 was detected at the ground surface.

The K-40 concentration values increase at 39.5 ft, gradually decrease from 40 to 62 ft, then gradually increase from 62 to 90 ft. A sharp increase in the U-238 concentrations was detected at 90 ft.

Additional information and interpretations of log data are included in the main body of the Tank Summary Data Reports for tanks C-105 and C-108.

Log Plot Notes:

Separate log plots show the man-made and the naturally occurring radionuclides. The natural radionuclides can be used for lithology interpretations. The headings of the plots identify the specific gamma rays used to calculate the concentrations.

Uncertainty bars on the plots show the statistical uncertainties for the measurements as 95-percent confidence intervals. Open circles on the plots give the MDL. The MDL of a radionuclide represents the lowest concentration at which positive identification of a gamma-ray peak is statistically defensible.

A combination plot includes the man-made and natural radionuclides, the total gamma derived from the spectral data, and the Tank Farms gross gamma log. The gross gamma plot displays the latest available digital data. No attempt has been made to adjust the depths of the gross gamma logs to coincide with the SGLS data.

An additional log plot compares spectral gamma data collected with the Radionuclide Logging System (RLS) in 1993 with spectral gamma data collected with the SGLS in 1997. Uncertainty bars and MDLs are not included on these plots.